


# Sentinel-3 Sea and Land Surface Temperature Radiometer Pre-Flight Calibration

Dave Smith  
STFC RAL Space

## Credits: SLSTR Core team




- Leonardo (formerly Selex ES), Instrument prime contractor, supply of Detector Assembly (the Focal Plane Assembly (FPA), the Front End electronics (FEE) and the Cryocooler (CCS)).
- JOP, supplier of opto-mechanical enclosure.
- RAL, responsible for calibration and systems design consultancy under ThalesAlenia as Sentinel 3A prime contractor.




# ATSR Series

## 1991-2000 ATSR-1



## 1995-2008 ATSR-2



## 2002-2012- AATSR



# SLSTR Series

2016 – Sentinel 3A



Launched 16-Feb-2016 😊

- 2017 – Sentinel 3B
- 2021 – Sentinel 3C
- 2023 – Sentinel-3D

...




Sentinel-3A First Image – 3-March 2016


# Key Requirements

- **Continuity of Sea and Land Surface Temperature datasets derived from (A)ATSR**
- **Additional bands for fire radiative power measurements and improved cloud detection**
- **Dual-View Capability**
- **On-board calibration sources**
- **Daily global coverage (with 2 satellites)**


AATSR Level-3 product at user-defined spatial resolution Europe daytime Feb 2011 at 0.25°



Global SST ENVISAT AATSR monthly composite




ENVISAT AATSR hot spot fires and world fire atlas



Casadio et al; ALGO3 persistent hot spot sites (1991-2009) RSE 2012



# SLSTR instrument

Nadir swath	>74° (1400km swath)
Dual view swath	49° (750 km)
Two telescopes	Φ110 mm / 800mm focal length
Spectral bands	TIR : 3.74μm, 10.85μm, 12μm SWIR : 1.38μm, 1.61μm, 2.25 μm VIS: 555nm, 659nm, 859nm
Spatial Resolution	1km at nadir for TIR, 0.5km for VIS/SWIR
Radiometric quality	NEΔT 30 mK (LWIR) – 50mK (MWIR) SNR 20 for VIS - SWIR
Radiometric accuracy	0.2K for IR channels 2% for Solar channels relative to Sun




# On-Board Calibration systems

## Thermal InfraRed Blackbodies



Effective  $e > 0.998$   
T non-uniformity  $< 0.02$  K  
T Abs. Accuracy  $0.07$  K  
T stability  $< 0.3$  mK/s  
8 PRT sensors + 32 Thermistors

## VIS-SWIR Channels VISCAL



Zenith diffuser +  
relay mirrors  
Uncertainty  $< 2\%$




# The Goal

To ensure the interoperability of satellite datasets it is a requirement for their measurements to be calibrated against standards that are traceable to SI units

For temperature this is the International Temperature Scale of 1990

For IR instruments such as SLSTR the traceability is achieved via internal BB sources




Instrument



Blackbody Source



S-PRT




Fixed Point Cells




# The Reality


## Calibration Sources




## Instrument Design



## Processing Model




## Calibration Facilities




## People



## Software



## Post Launch Activities



## Policy





## Processing Facilities




## VIS/SWIR Calibration

- SLSTR VIS/SWIR channels are calibrated via a diffuser based calibration VISCAL system – based on (A)ATSR concept
  - VISCAL is illuminated once per-orbit by the Sun
- Pre-Launch Calibration is to characterise key instrument performance
  - Radiometric response of each detector
  - Signal-to-Noise performance of each detector
  - Reflectance factor of VISCAL system
  - Polarisation sensitivity




## Source Setup

- Integrating sphere used for calibration of SLSTR
- 6 lamps, one (lamp 3) has a variable aperture. 0%=open, 100%=closed. Percentage is not proportional to open area.
- Lamp settings controlled and data recorded using labview interface on a PC




## Source Setup

- Three spectrometers mounted on the sphere to monitor source output and traceability to NPL calibration
  - 2 SWIR
  - 1 for VIS-NIR
- Lamp settings controlled and data recorded using labview interface on a PC



# Calibration of spectrometers

- Integrating Sphere calibrated at NPL against standard BB source August 2015



**Figure 1: Absolute Spectral Radiance of RAL U2000C Integrating Sphere measured in August-September 2015. Dashed lines show the  $k=2$  uncertainty of the measurements.**




# Uncertainty in Absolute Calibration


**Table 1: Breakdown of uncertainties in transfer of absolute radiance calibration to sphere spectrometers.**

Symbol	Component	Sensitivity Coefficient	Uncertainty Estimate	Divisor	Characterisation
$u(L_{\text{sphere}})$	Sphere calibration	1	<1% (VIS) <2% (SWIR)	2	NPL Report values quoted at $k=2$
$u(L_{\text{interp}})$	Interpolation of radiance	1	<0.2% (VIS) <2.5% (SWIR)	1	Comparison of interpolation methods - linear and quadratic. Depends on wavelength and presence of spectral features.
$u(\text{Noise\_All})$	Spectrometer Noise - full lamps	1	<0.29% (SWIR)	1	Standard deviation of signals during measurements
$u(\text{Noise\_dark})$	Spectrometer Noise - Dark Signal	1	<0.26% (SWIR)	1	Standard deviation of signals during measurements
$u(\text{Spec\_Drift})$	Spectrometer Drift	1	1%	$\sqrt{3}$	Comparison of spectrometer signals during calibration
$u(l)$	Spectrometer Wavelength	$\partial L/\partial l$ <1%/nm (VIS) <3%/nm (SWIR)	<1nm	1	Estimate should be calibrated with known spectral lines to get accurate wavelength registration

# Uncertainty in Absolute Calibration



# Processing of data from spectrometers



## Band Averaged Spectral Radiances:




are obtained by integrating the sphere radiance,  $L_\lambda$ , over the spectral response  $R_\lambda$ , at each band, using:

$$L_i = \frac{\int_0^\infty R(\lambda) I(\lambda) d\lambda}{\int_0^\infty R(\lambda) d\lambda}$$



# Spectrometer inter-comparisons

During the SLSTR testing (150mm, 75mm, & 50mm output port):






**S4:**  
Aperture 150 mm, Hamamatsu = NirQ  
Reduced Apertures: NirQ > Hamamatsu

**S6:**  
Aperture 150 mm, Hamamatsu ≥ NirQ  
Reduced Apertures: Hamamatsu > NirQ

From data measure



A & VIS13B):













# Radiometric Response – Nadir View



## SLSTR-A




## SLSTR-B




# SNR Performance

## SLSTR-A


S1 SNR




S2 SNR




S3 SNR




S4 SNR



S5 SNR




S6 SNR




## SLSTR-B


vis04:S1 SNR




vis04:S2 SNR




vis04:S3 SNR




vis04:S4 SNR



vis04:S5 SNR



vis04:S6 SNR



# Theoretical VISCAL reflectance factors




$$R_{VISCAL} = \tau_{UV} R_D(\theta_0, \theta_v, \varphi_0, \varphi_v) \rho_{m1} \rho_{m2} \rho_{m3} A_{m3} / A_{SLSTR}$$

NADIR	S1	S2	S3	S4	S5	S6
	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]
$R_D$	0.3257 [0.93]	0.3252 [0.94]	0.3229 [0.92]	0.3205 [2.66]	0.3218 [2.60]	0.3096 [3.01]
$\rho_{m1}$	97.27 [0.1]	98.13 [0.1]	98.16 [0.1]	98.62 [0.1]	98.78 [0.1]	99.21 [0.1]
$\rho_{m2}$	97.27 [0.1]	98.06 [0.1]	98.06 [0.1]	98.64 [0.1]	98.87 [0.1]	99.24 [0.1]
$\rho_{m3}$	97.33 [0.1]	98.03 [0.1]	97.97 [0.1]	98.16 [0.1]	98.32 [0.1]	98.62 [0.1]
$\tau_{UV}$	91.95 [0.14]	92.19 [0.13]	92.38 [0.46]	92.58 [0.42]	92.56 [0.43]	87.34 [0.5]
$A_{m3} / A_{SLSTR}$	0.21696051 [1.44]					
$R_{VISCAL}$	0.1880 [1.4]	0.1928 [1.4]	0.1917 [1.4]	0.1931 [2.1]	0.1949 [2.0]	0.1790 [2.1]

OBLIQUE	S1	S2	S3	S4	S5	S6
	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]	[3 $\sigma$ ]
$R_D$	0.3219 [0.98]	0.3211 [1.01]	0.3190 [1.00]	0.3172 [2.70]	0.3178 [2.62]	0.3060 [3.02]
$\rho_{m1}$	97.14 [0.1]	97.97 [0.1]	98.06 [0.1]	98.36 [0.1]	98.47 [0.1]	98.80 [0.1]
$\rho_{m2}$	97.32 [0.1]	98.06 [0.1]	98.03 [0.1]	98.48 [0.1]	98.58 [0.1]	98.84 [0.1]
$\rho_{m3}$	97.31 [0.1]	98.08 [0.1]	98.05 [0.1]	98.24 [0.1]	98.43 [0.1]	98.54 [0.1]
$\tau_{UV}$	91.95 [0.14]	92.19 [0.13]	92.38 [0.46]	92.58 [0.42]	92.56 [0.43]	87.34 [0.5]
$A_{m3} / A_{SLSTR}$	0.1869269 [1.77]					
$R_{VISCAL}$	0.1599 [1.6]	0.1638 [1.6]	0.1631 [1.6]	0.1641 [2.1]	0.1651 [2.1]	0.1510 [2.1]

Input data from component level characterisation – part of instrument level calibration database

# Measured Reflectance Factors



Measured SLSTR-A Nadir view Rcal for each detector (blue) compared to prediction (red)

Results show unexpected dispersion of measured reflectance factors.


Particularly pronounced in SWIR channels.

With exception of S6, average values in good agreement with predictions.

Similar results for Oblique view and SLSTR-B

# VISCAL Pixel Range and Uniformity

S3



We performed a set of measurements where the source illuminated the diffused and measured the signal response for different scanner positions.


Results determined the range of pixels to use on-orbit.

Showed a significant non-uniformity in the measured responses.


- For SWIR channels different for each detector
- Greater than expected variation in diffuser BRDF

**Why?**

S5a




S5b




# Pupil Uniformity – Along Scan

S3



To investigate cause of non uniformity we performed some additional measurements at centre of earth view.


S5a



We illuminate the earth view with a 50mm diameter source (i.e. underfilling the pupil) and measure the instrument response as a function of scanner position (along scan direction)

Results show all VIS channels appear to fill main aperture uniformly.


S5b



Differences seen in SWIR channel A and B stripes. Less uniform response

# Pupil Uniformity – Along Track


S3



We then repeated the measurements, this time moving source in vertical direction (along track direction)

Results show all VIS channels appear to fill main aperture uniformly.


S5a



Noticeable differences seen in each SWIR detector.

Conclusion:  
Main telescope aperture is not the primary pupil for the SWIR channels

S5b



Provides root cause for variations in measured instrument response and Rcal




# NPL-RAL-TAS Sphere Intercomparisons

An exercise has been initiated to compare spectral radiances of integrating sphere sources used for SLSTR (RAL Space) and OLCI (Thales Alenia Space, France) calibrations.

NPL are performing measurements using spectroradiometers and reference source at host institution.

Measurements performed at RAL in December during SLSTR calibration campaign. Data being processed.

Dates for OLCI to be confirmed - but close to instrument calibration




NPL's ASL spectrometer and source viewing RAL integrating sphere source.

## IR Instrument Calibration – Objectives

- **Does the end-to-end instrument calibration scheme work?**
  - New optical design – 2 telescopes not 1, multiple detectors per channel
  - OME thermal design – not based on AATSR heritage
- **Does the instrument calibration work over the full field of view and dynamic range?**
  - Wider instrument swath compared to AATSR
  - Nonlinearity, Noise performance, Dynamic range
- **Does calibration work in flight representative environment?**
  - Nominal BOL – is this defined?
  - EOL (Hot)
  - Orbital temperature variations

# Thermal IR Calibration Facility



Initial Trials with STM completed April 2012

TV and calibration of S3A instrument March-May 2015


S3B Calibration Oct 2016 – Feb 2017

S3C 2019

S3D 2020...

- ESA requirement to perform calibration tests under flight representative conditions.
  - *Thermal balance*
  - *Steady State*
  - *Instrument fully operational*


# TIR calibration- Blackbody Calibration Source



Precision RIRTs  
Calibrated to ITS90  
< 0.01K

Radiometric Accuracy  
< 0.05K

Emissivity  
 $12\mu\text{m} = 0.99871$   
 $11\mu\text{m} = 0.99870$   
 $3.7\mu\text{m} = 0.99911$



Sources  
previously  
used for all  
ATSR  
instruments

Standards



ATSR



ATSR-2



AATSR



**S3 SLSTR**


# IR Calibration Test Summary



- Calibration at 'Nominal' BOL conditions
  - Centre of Nadir/Oblique views
  - On-Board BBs at nominal settings (250K, 300K)
  - Test over full dynamic range (5K intervals)
  - Test over full swath (reduced number of scene temperatures)
- Calibration at 'Nominal' EOL conditions
  - Centre of Nadir/Oblique views
  - On-Board BBs at nominal settings (250K, 300K)
  - Test over part dynamic range (10K intervals)
- Tests with different on-board BB temperatures
  - Test performed at 'Nominal' BOL conditions
  - Currently at 'low', 'medium', 'high' power settings
  - +Y and -Y BBs will be switched
  - Test over part dynamic range (10K intervals)
- Orbital simulation tests

# Data Processing


- For SLSTR (as for ATSR series) we process data using calibration algorithms and input data as used by IPF.




# TIR Calibration - Measured vs Actual BT



Nadir




Oblique




# Measured - Actual BT SLSTR-A

Nadir



Oblique







# Measured - Actual BT SLSTR-B



## Nadir



## Oblique



## Calibration Model

- From the measured DN we wish to obtain the scene radiance  $L_{\text{scene}}$
  - Assuming that the radiometric response of the system is linear with radiance (or adjusted for detector non-linearity), we can derive the gain using two calibration sources of known scene radiance
    - i.e. Blackbodies
    - $DN_{\text{BB}} = g(L_{\text{BB}}) + DN_{\text{Offset}}$
  - This gives
    - $L_{\text{scene}} = XL_{\text{hbb}} + (1-X)L_{\text{cbb}}$
- Where
- $$X = (DN - DN_{\text{cbb}}) / (DN_{\text{hbb}} - DN_{\text{cbb}})$$
- **Both  $g$  and  $DN_{\text{offset}}$  MUST be constant during the calibration interval.**

## Stray light effects on calibration

- At TIR wavelengths specifically, the offset signal  $DN_{\text{offset}}$  is a combination of
  - Offset voltageAND
  - Thermal emissions from instrument (about 10-20% of the detected signal)
- A fundamental requirement for the calibration to work is that the thermal emission as viewed by the detector is constant for the full scan and over the calibration period (~10s).
- Stray light paths from other parts of the instrument should not cause a problem where they contribute to thermal background signal
  - Provided that they are NOT dependent on scan mirror position

## Scan Dependent strays

- What if  $DN_{\text{offset}}$  is not constant during the scan cycle?
- Lets consider as a function of pixel position, a scan dependent radiance perturbation of  $\pm\Delta L(\text{pos})$  which in turn gives rise to a perturbation in the background signal  $\pm\Delta DN(\text{pos})$ .

- The calibration model now becomes

$$L_{\text{scene}} + \Delta L(\text{scene}) = X(L_{\text{hbb}} + \Delta L(\text{hbb})) + (1-X)(L_{\text{cbb}} + \Delta L(\text{cbb}))$$

where

$$X = \frac{((DN_{\text{scene}} + \Delta DN(\text{scene})) - (DN_{\text{cbb}} + \Delta DN(\text{cbb})))}{(DN_{\text{hbb}} + \Delta DN(\text{hbb}) - (DN_{\text{cbb}} + \Delta DN(\text{cbb})))}$$

- But we are assuming the ideal calibration model, so the error we observe in the calibration is

$$\Delta L_{\text{error}} = \Delta L(\text{scene}) - X\Delta L(\text{hbb}) - (1-X)\Delta L(\text{cbb})$$

which can be rewritten as

$$\Delta L_{\text{error}} = (\Delta L(\text{scene}) - \Delta L(\text{cbb})) + X(\Delta L(\text{cbb}) - \Delta L(\text{hbb}))$$

## Error Model

Calibration Error Model due to offset errors is given by

$$\Delta L_{\text{error}} = \Delta L(\text{scene}) - X\Delta L(\text{hbb}) - (1-X)\Delta L(\text{cbb})$$

Where

$$X = (L_{\text{scene}} - L_{\text{CBB}}) / (L_{\text{hbb}} - L_{\text{cbb}})$$

Stray light model gives each term

$$\Delta L(\text{scene}) = f(L_{\text{stray,scene}} - L_{\text{scene}})$$

$$\Delta L(\text{cbb}) = g(L_{\text{stray,cbb}} - L_{\text{cbb}})$$

$$\Delta L(\text{hbb}) = g(L_{\text{stray,hbb}} - L_{\text{hbb}})$$

f and g are stray light factors derived from pre-launch calibration

$L_{\text{stray}}$  are derived from instrument temperatures (from instrument telemetry)

$L_{\text{cbb}}$ ,  $L_{\text{hbb}}$  are derived from blackbody thermometers (from instrument telemetry)





$L_{\text{scene}}$  is a function of temperature

Correction to BT is then

$$\text{BT}_{\text{corrected}} = \text{BT}_{\text{measured}} - \Delta T(\Delta L_{\text{error}})$$

# 'Stray Light' Model Correction



- Model provides good estimate measured calibration errors. Hence conclusion that this is best explanation for discrepancy.
- Input parameters derived from instrument temperatures available in HK.
  - These provide an approximation of the stray light source.
- Model has been coded and tested in Prototype Instrument Processing Facility (IPF-P)
- Early intercomparisons with IASI performed by EUMETSAT suggest that on-orbit stray-light error correction is not necessary.



## Conclusions

- Pre-launch calibration testing under flight representative is essential:
  - For TIR instruments this is particularly true since vicarious calibration – extremely challenging
  - Necessary for demonstrating end-to-end instrument calibration model
  - Allows identification, analysis and correction of measurement errors
  - Provides reference data that are needed for validating data processors and for post-launch activities
- Calibration testing takes time and resources to perform and process data in timely manner
  - As usual calibration is the last activity in an instrument build .... Huge pressure on schedule, budgets ... pressure to descope calibration activities.
- Maintaining a user perspective and objectivity are critical!
  - Vendor's focus is usually on meeting project requirements

# Thank You For Your Attention



## SLSTR L1b 10-Jul-2016